



www.servicecal.co.uk

Servicecal Ltd, Unit A9, Axis Point, Hareshill Business Park, Hilltop Road, Heywood, OL10 2RQ.
Tel : 01706 367008 Fax : 01706 622469 Email : info@servicecal.co.uk

Certificate of Calibration

Instrument Description : **Digital Multimeter**

Certificate Number : **261312**

Manufacturer : **Fluke**

Customer : **Waters Corporation**

Model Number : **34411A**

Customer Reference : **4070**

Serial Number : **MY48003234**

Order Number : **-**

Test Conditions :- Ambient Temperature 20C +/- 2C Humidity 50%RH +/- 10%RH

The above instrument has been calibrated with instruments that are traceable to National or International standards.
Traceability information is retained for each instrument used and copies are available on request.

This certificate is issued subject to our standard terms and conditions.

The uncertainty limits quoted refer to the measured values only at the time of calibration,
with no account being taken of the instruments ability to maintain its calibration.

The limits have been derived in accordance with 17025 and unless quoted otherwise are for a confidence
level of approximately 95%.

The results at the time of calibration for the above instrument were:-

Found not to meet specification

☐

Found to meet specification

☒

Quoted in absence of specification

☐

Found to meet spec after repair

☐

Found to meet spec after adjustment

☐

Further Remarks:

The instrument was externally cleaned prior to calibration.

At the points tested the instrument was found to meet manufacturer's specification.

APPROVED SIGNATORY:

M. Ahmad



DATE: 25th October 2017