

5414

Tektronix (China) Co. Ltd.
1227 Chuan Qiao Road
Pudong New District, Shanghai, China 201206

TRACEABLE CALIBRATION
TDS3054C, C020604
Cal Date: 09-MAY-2017
Cal Due: 19 DEC 2018 Lab: Tek
Tektronix

Tektronix**CERTIFICATE OF TRACEABLE CALIBRATION****Certificate No:** PC0KZ569FN**Revision:** 00**Manufacturer:** Tektronix**Model:** TDS3054C**Serial No:** C020604**Description:** Oscilloscope; 500MHz, 4 Channel**Temperature:** 25.5 °C**Humidity:** 47 %**Cal Date:** 09-MAY-2017**Date Placed In Service:** 19 DEC 17 ***Due Date:** 19 DEC 18 *

*Optional customer entry fields: The "Due Date" may be established by adding the calibration interval to the "Date Placed In Service." Recommended calibration intervals, for Tektronix products, are available on the Tektronix website at:
http://www.tek.com/mytek_support_status/

INSTRUMENT CONDITION:**Received:** Not applicable, initial calibration**Returned:** In tolerance, within published measurement specification

Tektronix certifies the above instrument has been calibrated using standards traceable to the PRC National Institute of Metrology (NIM) and/or other National Metrology Institutes (NIST, NPL, PTB) that are linked to the international system of units (SI). The policies and procedures used for the calibration of this product comply with ISO/IEC 17025:2005.

This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

CALIBRATION PROCEDURE:

MANIFEST:Product_Oscilloscopes_TDS3000-2012_Full VERSION:33

CALIBRATION EQUIPMENT USED:

<u>MANUFACTURER/MODEL</u>	<u>MODEL DESCRIPTION</u>	<u>ID NUMBER</u>	<u>DUE DATE</u>
Fluke 9500B	Oscilloscope Calibrator	CT0410	15-Oct-2017
Fluke 9530	Active Head, 9500	CT0308	15-Oct-2017
Fluke 9530	Active Head, 9500	CT0312	19-Nov-2017
Fluke 9530	Active Head, 9500	CT0386	15-Oct-2017
Fluke 9530	Active Head, 9500	CT0447	15-Oct-2017
Fluke 9530	Active Head, 9500	CT0483	15-Oct-2017

Issued By:**Certified By:** Zhongyuan Cao**Quality Director:** Cui, Hu Wa**Date Issued:** 11-MAY-2017

ISO 9001 Registered Quality System

TDS3054C C020604

Page 1 of 1

001-1387-01

