

Data Coverage : 01 May 2016 to 31 May 2018

Test Performed	Qty tested	Qty Passed	Qty Failed	% Fail
Degasser Test	9446	9431	15	0.159

Date	Model	Serial Number	FAIL Mode	Debug Findings	Debug Category	Results
5-May-16	186269505	D16SM7584A	Degasser test	Cannot detect degasser	Component faulty	FAIL
5-May-16	186269505	D16SM7584A	Degasser test	Cannot detect degasser	Component faulty	FAIL
9-May-16	186269506	E16SM4526A	Degasser test	Cannot detect degasser	Component faulty	FAIL
21-Nov-16	186269506	L16SM4308A	Degasser test	2.3 PSIA	Component faulty	FAIL
13-Dec-16	186269506	M16SM4452A	Degasser test	At pressure 3.4PSIA	Component faulty	FAIL
14-Dec-16	186269506	M16SM4451A	Degasser test	Degasser H/W fault (3) due to degasser chamber leak	Component faulty	FAIL
14-Dec-16	186269506	M16SM4490A	Degasser test	Degasser chamber leak	Component faulty	FAIL
15-Dec-16	186269506	M16SM4500A	Degasser test	Degasser H/W fault (3) due to degasser chamber leak	Component faulty	FAIL
7-Jan-17	186269506	A17SM4695A	Degasser test	Degasser pressure failed	Component faulty	FAIL
10-Jan-17	186269506	A17SM4700A	Degasser test	Degasser pressure failed	Component faulty	FAIL
9-Jan-17	186269506	A17SM4719A	Degasser test	Degasser pressure failed	Component faulty	FAIL
11-Jan-17	186269506	A17SM4738A	Degasser test	Degasser pressure failed	Component faulty	FAIL
7-Jul-17	186269505	G17SM7020A	Degasser test	No reading	Component faulty	FAIL
14-Sep-17	186269506	J17SM4970A	Degasser test	Degasser failed 1.6PSIA after 20 minutes boot up	Component faulty	FAIL
5-Dec-17	186269506	M17SM4738A	Degasser test	No reading	Component faulty	FAIL