
Waters Knowledge Base

Does the LC systems need qualification after Lace replacement? - wkb13804

ENVIRONMENT

- Lace
 - Empower
 - Waters HPLC, UPLC
-

ANSWER

- No need to qualify the systems, only the lace needs qualification.
 - Once new lace is installed, connectivity needs to be verified.
 - As a caution, please verify your internal SOP if it states otherwise
-

ADDITIONAL INFORMATION

