

|                          |                                  |
|--------------------------|----------------------------------|
| Instrument Type          | Beacon HTL and Source Baking Rig |
| Instrument Serial Number | TB726                            |

|                        |                                                                      |
|------------------------|----------------------------------------------------------------------|
| Correct Fuse(s) Fitted | <input checked="" type="radio"/> Yes <input type="radio"/> No / N/A* |
|------------------------|----------------------------------------------------------------------|

| Test Equipment | Control Number | Calibration Due |
|----------------|----------------|-----------------|
| PAT Tester     | 5119           | 03 JULY 2018    |
| Flash Tester   | 5119           | 03 JULY 2018    |

| Test Instruction 12252                           | Instrument PAT Test Results                                       |
|--------------------------------------------------|-------------------------------------------------------------------|
| Visual check on lead, and case of instrument     | <input checked="" type="radio"/> Pass <input type="radio"/> Fail* |
| Visual check on 13A plug, wiring and fuse rating | <input checked="" type="radio"/> Pass <input type="radio"/> Fail* |
| P.A.T Earth Bond Test Limit <0.1Ω                | 0.037 Ω                                                           |
| P.A.T Insulation Test Limit >2MΩ                 | 35.85 MΩ                                                          |
| P.A.T Leakage Test Limit <2.65mA                 | 1.27 mA                                                           |

| Outlet Earth Bond Results (if there are no outlets enter N/A) |       |       |       |       |
|---------------------------------------------------------------|-------|-------|-------|-------|
| P.A.T Earth Bond Test Limit <0.1Ω                             | N/A Ω | N/A Ω | N/A Ω | N/A Ω |

| Test Instruction 12252   | Instrument Flash Test Results |                                                                   |
|--------------------------|-------------------------------|-------------------------------------------------------------------|
| Flash test<br>Limit <1mA | Reported Value                | 0.02 mA / N/A*                                                    |
|                          | Overall result                | <input checked="" type="radio"/> Pass <input type="radio"/> Fail* |

|                              |                                                                   |
|------------------------------|-------------------------------------------------------------------|
| Overall PAT and Flash Result | <input checked="" type="radio"/> Pass <input type="radio"/> Fail* |
|------------------------------|-------------------------------------------------------------------|

Completed By:

| Printed Name | Signature                                                                           | Date        |
|--------------|-------------------------------------------------------------------------------------|-------------|
| A BARBER     |  | 28 MAR 2018 |

\*Circle as appropriate