

NOTE: Tenderers are to submit one performance summary for every offer quoted using the same format. You are advised to be truthful in your submission. Tenderer shall provide information where applicable. Tenderer shall specify "NA" if it is not applicable.

PERFORMANCE SUMMARY

Contractor	:	Waters Pacific Pte Ltd
Name of Unit/System	:	Acquity Arc/SQD2
Manufacturer/model	:	Waters Corporation
Year of manufacture	:	2016
Year of model	:	2015
Country of Origin	:	Singapore/UK
Warranty Period	:	1 year
Lead Time	:	4 – 6 weeks

No cross-referencing should be used in this performance summary unless specified by the Company.

- 1. SOFTWARE VERSION** Please specify
Empower 3 & Masslynx v4.2
- 2. OVERALL WEIGHT, Kg** Please specify
Please refer to Specifications Sheet
- 3. OVERALL SIZE, Height (H) x Width (W) x Depth (D)** please specify
Please refer to Specifications Sheet

4. POWER CONSUMPTION

- a. The individual power consumption of each and every sub-system tendered shall be indicated in Table 1. If special power requirement exist, they shall be highlighted.

Equipment Model	Voltage (V)	Phase	Standby Operation			Normal Operation		
			Current (A)	kVA	kW	Current (A)	kVA	kW
Quaternary Solvent Manager-R	100 to 240 V	Single	5.0A		0.36	5.0A		0.36
30-cm CHC Sample Manager	100 to 240 V	Single	3.0A		0.24	3.0A		0.24
FTNR	100 to 240 V	Single	10.0A		0.42	10.0A		0.42
2998 PDA Detector	100 to 240 V	Single	3.25A		0.195	3.25A		0.195

I hereby certify that all information are true and correct and shall fully comply with all requirements stated in this specification, unless stated otherwise under the Statement of Non-compliance.

Joanna



Signature & Stamp/Seal of the Tenderer

Joanna Ang / Account Manager
 Name/Designation

24/10/2016
 Date