

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
(1) Critical Specifications	
· The proposed system must be a Fast Liquid Chromatograph	Yes/ No
· The entire system must be a unified platform to be supplied by a single Tenderer to ensure seamless operation	Yes/ No
· Tenderer must be the direct manufacturer with an established network for after-sales support – engineers, application and logistics in Singapore	Yes/ No
· The Tenderer must demonstrate that it has a proven appropriate set-up and capability to provide after-sales service efficiently and effectively in Singapore for the proposed system in the tender or its similar series. Please provide the awarded contracts for the sales/ provision of maintenance support from 2016 to 2018 in Annex E Performance Summary.	Yes/ No
(2) Binary/Quaternary pump	
· Allow gradient formation	Yes/ No
· Should have at least two dual pistons in series pumps with servo-controlled variable stroke drive, floating piston design and active inlet valve (or equivalent)	Yes/ No
· Should be capable of standalone operation when software is not in use	Yes/ No
· Flow range: minimum 0.050 to 5.000 mL/min with 0.01 mL/min increments (or better)	Yes/ No
· Flow precision: $\leq 0.075\%$ RSD	Yes/ No
· Flow rate accuracy: $\pm 1\%$ of set value or 10 mL/min, whichever greater	Yes/ No
· Operating pressure range: At least 0 – 800 bar for the entire range of operating flow rates	Yes/ No
· Pressure pulsation: $\leq 2\%$ amplitude or ≤ 3 bar or whichever better, at 1 ml/min isopropanol for all pressures above 10 bars	Yes/ No
· Compressibility: Pre-defined, based on mobile phase compressibility	Yes/ No
· Working pH range for all components: minimum 2 to 12 to be available as standard	Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
· Compatible with phosphate buffers.	Yes/ No
· Pump delay volume: ≤ 900 mL	Yes/ No
· Composition range: Settable from 0 – 100% per channel	Yes/ No
· Composition precision: ≤0.5% RSD	Yes/ No
· Composition accuracy: ≤ 0.5% absolute or less	Yes/ No
· Should have solvent selection valve in order to combine 2 out of 4 solvents for binary gradient or to select different solvent for flushing column	Yes/ No
· Pump must be able to run with existing columns sizes as follow: • ID 2.1 mm, 3 mm and 4.6 mm; • 5 µm, 3.5 µm and sub-2-µm particle columns; and • Superficially porous columns.	Yes/ No
· Inclusive of solvent cabinet; complete with minimum five sets of filters and caps (with 4 holes) and automatic rinsing kit	Yes/ No
· Tenderer is required to provide solvent cabinet(s) that able to hold at least one 5-litre round bottle (with circumference of approx. 60cm) together with three 1-litre round bottles (with circumference of approx. 33cm)	Yes/ No
(3) Solvent degasser	
· No. of independent channels: at least 4	Yes/ No
· Internal volume: At least 400 uL per channel to allow for fast and easy flushing, purging, priming, and solvent exchange	Yes/ No
· Working pH range for all components: at least 2 – 12 as standard	Yes/ No
(4) Auto-sampler	
· Operating pressure range should be able to reach up to at least 800 bars	Yes/ No
· Able to accommodate minimum 96 samples of 2 mL vials	Yes/ No
· Able to use vials with rubber, silicone and PTFE septa	Yes/ No
· Minimum Injection volume: 0.1 - 10 mL without loop change in ≤ 0.1 mL increments	Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
· Precision: < 0.3% RSD	Yes/ No
· Accuracy: ≤2%	Yes/ No
· Automated or programmed needle wash function which can be turned on or off to minimise carryover. Needle wash function should preferably not impact sample loop and injection volume. Tenderer should provide clear information from existing published material on how the needle wash is carried out	Yes/ No
· Additional needle flush station with a peristaltic pump or equivalent function to wash the exterior of the needle	Yes/ No
· Injection cycle time: Not more than 50 sec per sample	Yes/ No
· Equipped with high pressure valve that allows use of 1.8-µm fully porous particle column or superficially porous column	Yes/ No
· Pre-treatment function that minimally includes reagent addition, dilution and mixing must be available.	Yes/ No
· Overlapped injection ability to increase productivity that can be used with both full and partial sequence run	Yes/ No
· Instrument must have a transport mechanism that uses an X-Z-Theta robot (or equivalent) to optimise vial positioning by the gripper arm of the autosampler.	Yes/ No
· Autosampler should be able to perform injection from the correct vial position consistently for all samples on sample tray. There should be no instances for wrong vial injected or skipping of vials.	Yes/ No
· Needle material or needle coating should be compatible with vials using rubber, silicone and PTFE septa and be able to withstand high throughput injections without significant wear off or carryover. Published information about the needle material and ruggedness should be submitted upon request.)	Yes/ No
· The price quoted shall include at least: (1) 1 x minimum 96-vial (2-ml) thermostatted tray (4°C to 40°C); and (2) 1 x 40-vial (2-ml) tray (if available)	Yes/ No
(5) Column compartment	
· Temperature range: 10 °C below ambient to at least 80 °C	Yes/ No
· Temperature accuracy: ± 1°C or less	Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
Column capacity: at least 2 columns of 300 mm length	Yes/ No
Option to purchase: Column switcher to switch minimally between two columns. Tenderer shall quote this item separately in the format set out in Part 3 Annex D – Price Schedule.	Yes/ No
(6) Diode Array Detector	
Detection type: At least 512-element photodiode array	Yes/ No
Lamp type: Deuterium lamp	Yes/ No
Able to monitor at least 4 compound specific wavelengths simultaneously during a single analysis without loss in sensitivity	Yes/ No
Signal and spectra sampling rate: up to at least 150 Hz	Yes/ No
Short-term noise: $\leq \pm 8 \mu\text{AU/cm}$ with 10 mm path length cartridge cell	Yes/ No
Signal drift: $< 1 \times 10^{-3} \text{ AU/hr}$ at 230 nm or equivalent	Yes/ No
Upper linearity limit: $\geq 2.0 \text{ AU}$ (Based on ASTM method)	Yes/ No
Wavelength range: At least 190 – 700 nm	Yes/ No
Wavelength accuracy: Better than $\pm 1.0 \text{ nm}$, self-calibration with deuterium lines, verification with holmium oxide filter	Yes/ No
At least two programmable slit width between 1 nm-8 nm	Yes/ No
- Include flow cell with equivalent or better features: 10 mm pathlength max, $< 2 \mu\text{L}$ volume, withstand at least up to 80 bar pressure	Yes/ No
Include data analysis software for spectra evaluation with purity and spectra matching functions. Tenderer to provide sufficient information and training to help users understand the peak purity function as well as the basis of spectra matching on the instrument.	Yes/ No
Able to create and register UV spectra to the UV library in the system. Please provide screenshot for the UV library	Yes/ No
Radio frequency identification (RFID) tags for flow cells and UV lamp that record parameters such as cell dimensions, lamp usage, serial number, etc.	Yes/ No
Able to select reference wavelength for elimination of background interference	Yes/ No

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
The system must have a built-in data recovery feature which help prevent data losses in the event of communication breakdown between instrument and PC	Yes/ No
(7) Chromatographic software	
The software should have an user interface that allows the complete control of the liquid chromatographic system including access to all method parameters, data acquisition, data analysis, method automation, reporting, generation of customised reports, and system management utilities	Yes/ No
Operating software should have multi-tasking capability – capable of control, data acquisition, data evaluation, automation and reporting.	Yes/ No
Software should minimally have functions for wavelength optimisation, generation of peak purity reports, spectral libraries search for compound identification, and mathematical operations such as normalisation, subtraction and smoothening.	Yes/ No
Must be able to run at the same time with other programs on the computer such as word processors and also export information to other programs through industry standard protocols and file formats	Yes/ No
Software should preferably have functions for updating of volumes for all four solvents attached to the pump(s) independently.	Yes/ No
Software should allow shutdown after a sequence is completed. Software should preferably also allow shutdown when calculated remaining solvent volume in any of the solvent bottles reaches below a certain threshold.	Yes/ No
Must have built-in or integration path to OpenLab ECM for central secure long-term data storage	Yes/ No
Configurable access control to different functions of software	Yes/ No
Software must be able to provide log/ trails for method and results changes.	Yes/ No
Reports are customisable according to the laboratory's needs. Assistance is to be provided by the Contractor to ensure the seamless integration and interface with LISA for the instrument report printing and extraction	Yes/ No
Software should preferably be able to read UV library in “.uvl” format. If the software to be provided is unable to read “.uvl” format, Contractor is required to provide assistance to ensure the library can be accessed and used on the new software.	Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
· Include software with intuitive diagnostic and monitor capabilities and alert functions for problem notifications	Yes/ No
· Software should preferably be able to include both chromatogram and purity results for each analysis in a single report template which can be printed out directly from the software	Yes/ No
(8) Computer and Printer	
· PC (Specifications of the following should be equivalent or better) • Intel® Core™ 7th Gen i7 Processor (3.6 GHz, 8 MB Cache, Quad core) • Monitor: 20" IPS LED Monitor • Hard Disk: 1 TB 7200 rpm SATA 6Gb/s HDD • Memory: 16 GB DDR4 RAM • Optical Mouse • Keyboard • Operating System: Microsoft Windows 10 Professional SP1 or newer version with Microsoft Office 2016 or newer version • 2 network cards	Yes/ No
· Printer: Laserjet network printer with at least 1200x1200dpi resolution and prints at least 32 pages per minute. The printer must be capable of automatic duplex printing.	Yes/ No
· Tenderer shall ensure that the instrument PC/notebook provided with the instrument is network ready and able to connect to HSA's network for interfacing with LISA (Laboratory Information & Scientific Administration) system. If required, additional network cards shall be included to ensure the connectivity to LISA	Yes/ No
(9) Other Features	
· Instrument and system should be compatible with existing LC methods – allow direct method transfer from current Agilent 1100 Series LC systems and Shimadzu Nexera Series UHPLC systems	Yes/ No
· Modular design that has flexible configuration options and true front accessibility for all components	Yes/ No
· Able to run in 2 modes: HPLC and fast LC without the need for modification of hardware	Yes/ No
· The proposed system must be incorporated with GLP features that includes:	Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
I) Early maintenance feedback with display for a) Pump: litres of mobile phase pumped, mobile phases and seal wear b) Autosampler: number of injections, user-settable limits messages c) Detector: lamp time, lamp energy, date of last lamp replacement II) User-set limits for feedback messages for timely replacement of parts III) Electronic records of maintenance and errors for all modules in a log book	
· The system shall come with safety and maintenance requirements that includes: Extensive diagnostics, error detection and display, leak detection, safe leak handling, leak output signal for shutdown of pumping system. Low voltages in major maintenance areas	Yes/ No
· There shall have storage of all details of analytical method including instrument set points, data acquisition, data evaluation and reporting parameters as one analytical specification.	Yes/ No
(11) Integration with HSA's LISA (Laboratory Information & Scientific Administration) System	
· Tenderer shall ensure the Instrument PC/notebook provided will be able to run (a) OpenLAB – Cerity Enterprise Content Manager (ECM) Version 3.3.2 (b) OpenLAB – Cyberprinter (c) Must have built-in or integration path to OpenLAB ECM for central secure long-term data storage	Yes/ No
(12) Middle-Tier Software for integration with LISA	
· Tenderer shall ensure the instrument software will be able to have direct interface to LISA for the following: • Raw Data Files archival & retrieval • Instrument test result report output to PDF for archival • Instrument test result report for result uploading.	Yes/ No
· Tenderer shall specify the format of the instrument software and outputs.	Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
a) Software: _____ b) Output format: _____	
· Tenderer shall propose a complete and comprehensive solution to ensure the seamless interface between the instrument software and LISA, and to meet the requirements as specified above. The proposal shall include required hardware, system software, application software and professional services.	Yes/ No
(13) Configuration and setting of Instrument PCs/Note for integration with LISA	
· Contractor shall work with HSA assigned personnel for the setting up and configuration of Instrument PCs / notebook to ensure the connection to the HSA network as well as integration to LISA	Yes/ No
· Contractor shall work with HSA personnel for the following: • To define and assign a Hostname and/or PC name given by HSA to facilitate the instrument management on HSA network • To define the Raw Data file structure & directory assigned by HSA to facilitate the automation of electronic archival and retrieval of files • To work with LISA vendor, if required, to ensure the instrument report print out to be in PDF file format • To perform UAT to ensure the integration to LISA is running fine	Yes/ No
(14) Instrument Reports Requirement	
· Contractor shall work with Laboratory to ensure the seamless integration and interface with the Laboratory's Information Management System for the instrument report printing and extraction	Yes/ No
· Tenderer shall provide HSA the following output format for the Instrument Test Result output and they are: • Raw Data File: _____ • Electronic output file: _____ • Hardcopy print out: _____ Please state if there are other output files can be generated by Instrument: _____	Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
· The following information must be able to be captured and printed on the above mentioned Instrument Test Result output format to facilitate the integration and interface for result extraction and they are: • Data file name • Method used • Operator • Date Acquired • Instrument • Sample Name • Misc Info • Vial Number	Yes/ No
The sequence and sequence printout should include the following information: • Sequence line • Type of sample • Vial number • Data file name • Name of Method • Sample Information (with at least 2 lines of information) • Amount of sample • Volume of sample • Dilution factor • Internal Standard Amount	Yes/ No
The library and library matching printout should include the following information: • Sequence line • Type of sample • Vial number • Data file name • Name of Method • Display of UV spectrum of target • Results of the UV library match	Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
Contractor shall assist user to configure the required information should additional information/ parameters need to be included.	Yes/ No
(15) Installation, Commissioning, Training and Warranty	
All systems proposed must be GLP compliant	Yes/ No
Prices quoted must include installation and commissioning, warranty, operator training and LC starter kit.	Yes/ No
Hands-on training for operating the equipment model submitted in this tender must be provided for 3-6 persons.	Yes/ No
An endorsed copy of 5x5 risk assessment form for the installation of the instrument must be submitted to the laboratory 2 weeks before the delivery. A template of this form can be found on MOM website.	Yes/ No
The system should come with at least three (3) years warranty that covers the system hardware, software and computer. The warranty shall minimally include the following: a) unlimited breakdown calls and on-site support services during Office Hours from 8:00 am to 5:30 pm Monday to Friday excluding Saturday, Sunday and Public Holidays; b) 2 preventive maintenance (PM) yearly; c) Supply of all replacement parts (non-consumable parts); and d) Provision of software upgrade or revisions. · Contractor should accommodate observation by laboratory staff during the maintenance process. Any software revisions should also be provided free-of-charge.	Yes/ No
(16) Technical support	
· Contractor shall provide full assistance in terms of hardware, software and application support. This is to ensure that the instrument is fully validated with relevant methods for casework within 3 months after commissioning	Yes/ No
(17) Delivery	
· The delivery of the equipment is to be coordinated with the Laboratory.	Yes/ No
(18) Option to purchase: 4-year post warranty comprehensive maintenance services on biennial renewal basis	

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>
Tenderer shall quote for the option to purchase 4-year comprehensive maintenance contract on biennial renewal basis. The contract shall minimally include the following: a) Unlimited breakdown calls and on-site service support during Office Hours; b) 2 PM per year; c) Supply of all replacement parts (non-consumable parts); and d) Provision of software upgrade or revisions The coverage for post warranty comprehensive maintenance contract must be extended to system hardware and software.	Yes/ No
Replacement and Purchase of Components - The Contractor shall inform the Authority of any part or component of the equipment that needs to be replaced. - The Contractor shall use all reasonable endeavours to supply only original or genuine replacement components, which are required to maintain the equipment in good working order. Guarantee - The Contractor hereby warrants that all the Equipment and every part or unit thereof serviced by the Contractor shall conform to the requirements of this Contract and be free from any defect or deficiency in their performance or workmanship for a period of three (3 months) from the date of acceptance of the Services by the Authority (hereinafter referred to as "the Equipment Guarantee Period"). The Contractor hereby also warrants that all materials and every part or unit thereof, used for the Services shall be free from any defect in manufacture or any defect arising out of design, workmanship or materials used or any failure or malfunction due to quality of material, workmanship, design, assembly, installation or integration for a period of three (3) months from the date of acceptance of the materials by the Authority (hereinafter referred to as "the Materials Guarantee Period"). The Equipment Guarantee Period and the Materials Guarantee Period shall collectively be known as the "Guarantee Period". - Where during the Guarantee Period, any part or unit of the Equipment serviced by the Contractor or materials used for the Services or supplied therefor are found to be defective of any of the reasons specific in Clause 3 above, the Contractor shall at the written request of the Authority replace the same at the expense (including freight charges, transportation costs or other incidental charges) of the	

Equipment: Fast Liquid Chromatograph Location: IDL	Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>		
Contractor or if the Authority so agrees make good or repair the same at the expense (including freight charges, transportation costs and other incidental charges) within a Guarantee Turn-Around-Time of 7 days. The Equipment or materials or part or unit thereof so replaced or repaired shall be subject to the Guarantee under this Clause for the Authority or three (3) months whichever is later and such balance of the Guarantee Period shall begin to run only from the date of receipt of the part or unit so replaced or repaired by the Authority.			
(19) Requirements for Sample Testing			
- Tenderer is required to perform testing for three (3) sets of samples (Sample A, B & C) and these must be tested without further manipulation and using the conditions provided in the <i>Instructions for Sample Testing</i> that will be provided during collection of Sample A. The samples must be analysed on <u>the equipment model proposed</u> in this tender using: a) Column: ZORBAX SB C8, 50 x 4.6mm i.d., 1.8 µm (Agilent Part No. 827975-906-SB-C8); and b) In-line Filter: Fast LC In-line Filter, 4.6mm, 0.2um filter (Agilent Part No. HP 5067-1553) - Only one (1) set of sample will be provided to Tenderer for the testing. - Annex F Summary Sheet for Sample Testing Results must be completed after the analysis is completed. Please see table below for the sample testing timeline & requirements.			
Set of Sample	A	B	C
Sample Collection	Tenderer shall collect sample between 9am to 12pm on or before 9 th Nov 2018 at: Address: 11 Outram Road, Singapore 169078 Contact number: 6213 0712	Samples will be sent down by laboratory staff and the analysis shall be performed on the same day approximately within 2 weeks after the tender is closed. Please indicate the availability of the proposed system to perform the testing between 16 th Nov 2018 (tender closing date) to 30 th Nov 2018.: _____	

Yes/ No

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

Equipment: Fast Liquid Chromatograph Location: IDL			Compliances (Yes/ No) <i>**If No, please indicate reason of non-compliances.</i>	
Submission of Annex F	To submit by tender closing date and time.	To submit by 7 th Dec 2018.		
Requirement for Sample Analysis Result	i) Carryover must be ≤ 0.1% without needle wash (if applicable); and ii) < 0.05% carryover with external needle wash; and iii) No carryover should be observed using 12 mg/mL of caffeine at UV wavelength of 240 nm with needle wash.	Tenderer should demonstrate that acceptable peak purity results (i.e. peak purities passed) can be obtained for standard solutions during evaluation with or without further data correction. The criteria of an acceptable peak purity should be indicated.	Tenderer should demonstrate that target analytes are identified through UV library matching with or without post processing during evaluation. Criteria of a positive identification by the software should be indicated.	
- The unutilized samples must be returned to the Authority after the test analysis is completed. - If Tenderer requires to extend the deadline for result submission, please write in to the Authority and a maximum of 7 working days can be provided. - The detail results and supporting documents stated in Annex F for Sample A shall be submitted by tender closing on 16th Nov 2018 and for Sample B & C by 7th Dec 2018.				

PART 3: REQUIREMENT SPECIFICATIONS**HSA000ETT18300072****Equipment: Fast Liquid Chromatograph****Location: Illicit Drugs Laboratory****Quantity: 2 unit**

COMPLIANCE TO REQUIREMENT SPECIFICATIONS

I hereby confirm that I can meet all the
Requirement Specifications as stated above.

YES / NO*

*Please indicate any non-compliance here should there be any.

Reference To	Remarks