

EVALUATION CRITERIA

The proposals will be evaluated on all the following criteria:

Critical Evaluation Criteria

Failure in fulfilling this criterion/these criteria will be disqualified from further evaluation.

- 1) Not debarred by the Ministry of Finance on or after the closing date;
- 2) Certification with Expenditure and Procurement Policies Unit ("EPPU") under Supply Head EPU/LAB/30 or EPU/LTE/10 with Financial Registration Grade S5 (up to S\$1 Million) and above or equivalent;
- 3) Full Compliance with Conditions of Contract; and
- 4) Full Compliance with Requirement Specifications

Evaluation Criteria

In addition to the above critical criterion/criteria, participating Tenderers will be evaluated based on the following criterion/criteria:

Equipment 1: LC-MS/DAD

- a) Rates and prices proposed in the tender – 60%
- b) Quality – 40%

Evaluation Criteria	Weight Assigned	Description
(a) Price	60%	$\frac{\text{Lowest Tender Price}}{\text{Tenderer's Price}} \times \text{Price Weightage (60\%)}$
(b) Non-Price - Quality	40%	(1) Linearity, R2 (5%) (2) Accuracy in determining the concentration in sample solution (20%) (3) Signal-to-noise (S/N) ratio (5%) (4) Other criteria (10%) (a) Clause 7(ii) - System is able to switch on and off for power saving, and be able to pump down and be ready for sample analysis within one hour after switching on for efficient use of instrument. (b) Clause 9(iv) – Able to extract UV data from MS chromatogram in the same window and vice versa. (c) Clause 9(v) - System is able to show real-time spectra and signal even when sample injection is not performed.

Equipment 2: GCMS

- a) Rates and prices proposed in the tender – 70%
- b) Quality – 30%

Evaluation Criteria	Weight Assigned	Description
(a) Price	70%	$\frac{\text{Lowest Tender Price}}{\text{Tenderer's Price}} \times \text{Price Weightage (70\%)}$
(b) Non-Price - Quality	30%	Minimum number of steps required for obtaining the library match during manual processing. Note: Not looking at batch processing.