

	Description	Requirement	Frequency of replacement
1	Suction Filter	Clean	Every 5-7 mths
		Replace	*
2	Lamp a) UV Lamp b) Tungsten Lamp	Replace	Not more than 2000 hrs
			Not more than 2000 hrs
3	Line Filter	Clean	Every 5-7 mths
		Replace	*
4	Check-In Valves	Clean	Every 5-7 mths
		Replace	*
5	Check-Out Valves	Clean	Every 5-7 mths
		Replace	*
6	Pump Heads	Clean	Every 5-7 mths
7	Pump A Plunger Seal	Replace	Every 5-7 mths
8	Pump B Plunger Seal	Replace	Yearly
9	High Pressure Stator	Clean	Every 5-7 mths
		Replace	*
10	Low Pressure Stator	Clean	Every 5-7 mths
		Replace	*
11	High Pressure Rotor	Clean	Every 5-7 mths
		Replace	Yearly
12	Low Pressure Rotor	Clean	Every 5-7 mths
		Replace	*
13	Needle Seal	Clean	Every 5-7 mths
		Replace	Yearly

* When necessary