



Part Number: 186269505
 Tested By: CHONG WOOLKHIP
 Serial Number: F18SM7171A
 UUT Firmware: 3.04

Description: 2695 Separations Module
 Test Template Revision: 1.8
 Test Station: Final-test-2
 System: 2695_1



	Date Acquired	SampleName	Area	Percent
1	6/23/2018 12:27:38 PM SGT	Pre Blank		
2	6/23/2018 12:30:22 PM SGT	Standard	11985	0.0100
3	6/23/2018 12:33:04 PM SGT	Challenge	2137152	
4	6/23/2018 12:50:12 PM SGT	Post Blank	457	0.0004

Carryover Test:
 PreBlank injection (to establish system cleanliness)
 Standard (0.4 PPM) Caffeine injection
 Challenge injection (10,000 X Standard = 4,000 PPM)
 PostBlank injection
 Carryover (%) = Carryover Area/Standard Area * 0.010
 Specification: Post Blank Percent < 0.002
 (or Blank peak(s) are missing/blank)

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests fail, a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing; the initial form must be attached and retained.