

Waters Corporation - Test Record

Part Number: 186248900Product: 2489 UV/Visible DetectorOperator: SU BINGJIEStart Time: 05/12/2018, 11:55:Serial Number: E1887E574AEnd Time: 05/12/2018, 12:55:Test System: 2489 Final Test Station TE82903-6

Name	LSL	Measurement	USL	Unit	Outcome
Drift	-0.0001	0	0.0001	AU	PASS
Drift Count	0	0	99	EA	PASS
Noise	0	0.0000029	0.000005	AU	PASS
Noise Count	0	1	99	EA	PASS
Dual Noise	0	0.000006	0.000035	AU	PASS
Dual Noise Count	0	1	99	EA	PASS
Linearity	0	3.98940283991301	5	%RSD	PASS
Wavelength Accuracy	254	255	256	nm	PASS
Wavelength Accuracy	377.8	378.4	379.8	nm	PASS
Wavelength Accuracy	521.4	522.7	523.4	nm	PASS

2489 Final Test Overall Result - PASS

Signed:

Date:

END OF DATA