

Waters Corporation - Test Record

Part Number: <u>186299800</u>	Product: <u>2998 refresh</u>
Operator: <u>246112</u>	Start Time: <u>08/02/2018, 17:06:</u>
Serial Number: <u>H18998939A</u>	End Time: <u>08/02/2018, 19:59:</u>
Test System: <u>TE82882-8</u>	Order: <u>NA</u>
Test Software: <u>4.0</u>	Location: <u>Flextronics,SG</u>

TMD: NA

Firmware: 1.70 Build#: 6363 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Electrical Safety		true			PASS
Lambda1	178.0	185.7189	188.0	nM	PASS
Lambda512	801.5	803.783	811.5	nM	PASS
Shunt/Cell energy 6ms	80000	192997	250000	counts	PASS
Total Runs	0	11	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0.0	3.64	8.0	uAu	PASS
Final Drift	-1000.0	963.0	1000.0	uAu	PASS

2998 Refresh Final Test Overall Result - PASS

Signed:	Date:
---------	-------

END OF DATA