

Waters Corporation - Test Record

Part Number: <u>186002998</u>	Product: <u>2998 PDA</u>
Operator: <u>978611</u>	Start Time: <u>07/28/2018, 14:01:</u>
Serial Number: <u>G18998919A</u>	End Time: <u>07/28/2018, 17:39:</u>
Test System: <u>TE82882-8</u>	Order: <u>NA</u>
Test Software: <u>4.0</u>	Location: <u>Flextronics,SG</u>

TMD: NA

Firmware: 1.70 Build#: 6363 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Electrical Safety		true			PASS
Lambda1	178.0	186.2816	188.0	nM	PASS
Lambda512	801.5	803.3029	811.5	nM	PASS
Shunt/Cell energy 6ms	80000	183036	250000	counts	PASS
Total Runs	0	14	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0.0	4.75	8.0	uAu	PASS
Final Drift	-1000.0	963.0	1000.0	uAu	PASS

2998 PDA Final Test Overall Result - PASS

Signed:

Date:

END OF DATA