

Waters Corporation - Test Record

Part Number: <u>186015033</u>	Product: <u>eLambda PDA</u>
Operator: <u>978611</u>	Start Time: <u>07/28/2018, 13:06:</u>
Serial Number: <u>G18UPL107A</u>	End Time: <u>07/28/2018, 13:33:</u>
Test System: <u>TE82882-10</u>	Order: <u>NA</u>
Test Software: <u>4.0</u>	Location: <u>Flextronics,SG</u>

TMD: NA

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Baseline Temperature (Audit)	20.0	26.0	40.0	Degrees C	PASS
Hot Baseline (Audit)	20.0	26.44	40.0	Degrees C	PASS
F1 alone (Audit)	20.0	27.85	40.0	Degrees C	PASS
F1 and F2 (Audit)	20.0	25.74	40.0	Degrees C	PASS
Electrical Safety		true			PASS
Lambda1	178.0	185.7291	188.0	nM	PASS
Lambda512	801.5	804.7306	811.5	nM	PASS
Shunt/Cell energy 6ms	80000	140450	250000	counts	PASS
Total Runs	0	1	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0.0	6.61	7.0	uAu	PASS
Final Drift	-1000.0	1000.0	1000.0	uAu	PASS

eLambda PDA Final Test Overall Result - PASS

Signed:

Date:

END OF DATA