

Waters Corporation - Test Record

Part Number: 186002998Product: 2998 PDAOperator: 978611Start Time: 07/28/2018, 14:57:Serial Number: G18998918AEnd Time: 07/28/2018, 18:50:Test System: TE82882-8Order: NATest Software: 4.0Location: Flextronics,SG

TMD: NA

Firmware: 1.70 Build#: 6363 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Electrical Safety		true			PASS
Lambda1	178.0	186.3757	188.0	nM	PASS
Lambda512	801.5	804.2207	811.5	nM	PASS
Shunt/Cell energy 6ms	80000	168739	250000	counts	PASS
Total Runs	0	15	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0.0	5.09	8.0	uAu	PASS
Final Drift	-1000.0	919.0	1000.0	uAu	PASS

2998 PDA Final Test Overall Result - PASS

Signed:

Date:

END OF DATA