

Waters Corporation - Test Record

Part Number: 186002998Product: 2998 PDAOperator: 246112Start Time: 08/07/2018, 11:37:Serial Number: H18998962AEnd Time: 08/07/2018, 12:27:Test System: TE82882-8Order: NATest Software: 4.0Location: Flextronics,SG

TMD: NA

Firmware: 1.70 Build#: 6363 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Electrical Safety		true			PASS
Lambda1	178.0	186.2904	188.0	nM	PASS
Lambda512	801.5	803.389	811.5	nM	PASS
Shunt/Cell energy 6ms	80000	158655	250000	counts	PASS
Total Runs	0	3	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0.0	6.06	8.0	uAu	PASS
Final Drift	-1000.0	898.0	1000.0	uAu	PASS

2998 PDA Final Test Overall Result - PASS

Signed:

Date:

END OF DATA