



Calibration / Preventive Maintenance Record
For
WATERS Test Equipment

(1) PERFORM DATE: 12 / 06 / 20 (2) DUE DATE: 14 / 06 / 20 (3) FREQUENCY: 12 months

(4) NEXT PERFORM DATE: 12 / 06 / 21 (5) DONE BY: 246394 f
(Badge/SIGN)

(6) CALIBRATION: YES ☒ (NO) (CIRCLE ONE)

(7) PREVENTIVE MAINTENANCE: ☒ YES ☐ NO (CIRCLE ONE)

(8) T.E. TESTER NO : TEB2943-11

(9) SYSTEM NAME : Bsm lqsm Final Test

(10) TESTER NAME : TEB2943-11, Acuity SDS Tester

(11) LOCATION : 5th Floor, Kalang Place (Production area)

(12) ADJUSTMENT MADE: YES ☒ (NO)
(ADJUSTMENT MADE DURING CALIBRATION / PREVENTIVE MAINTENANCE)

(13) EQUIPMENT USED	SERIAL NO	CAL DATE	CAL DUE DATE
DC Tracking Power Supply	7090053	18 / 05 / 20	18 / 05 / 21
Digital multimeter	90240092	13 / 02 / 20	13 / 02 / 21
	End of data	/ /	/ /
		/ /	/ /
		/ /	/ /

(14) Procedure Step #	Expected	Actual	Remarks
1.7	0.000 $\pm$ 0.022 VDC	0.002 VDC	NA
1.7	0.000 $\pm$ 0.022 VDC	-0.001 VDC	NA
1.9	2.000 $\pm$ 0.022 VDC	2.003 VDC	NA
1.9	2.000 $\pm$ 0.022 VDC	2.002 VDC	NA

NOTES:

(15) EQID NO: PM 02477

(16) REPORT NO: NA

(17) PROCEDURE NO: WAT-RW-SKI01716 (735001828) Rev B

(18) REVIEWED BY: JOEL MAMALATEDO DEYUZA T 12/06/20
NAME/SIGN/DATE (DD/MM/YY)

* correction of wrong entry

f mohd shamsul 246394 12 Jun 20