


First Pass Yield Histogram for All Reports and Modes [P] on
[PDA Detector Final Test TE82882-10, PDA Detector Final Test TE82882-11, PDA Detector Final Test TE82882-13, PDA Detector Final Test
TE82882-14, PDA Detector Final Test TE82882-8, PDA Detector Final Test TE82882-9] @ Detectors @ LC Detectors
Time Period - 03/30/2020 00:00:00 HKT to 05/31/2020 23:59:59 HKT



26-May @ 10:30 AM

Legend

Sl	Item	Pass	Fail	Total	First Pass Yield [%]
04	All Reports and Modes [P] on LC Detectors/Test Station/Station/PDA Detector Final Test TE82882-8	18	2	20	90
05	All Reports and Modes [P] on LC Detectors/Test Station/Station/PDA Detector Final Test TE82882-9	53	11	64	82.8
00	All Reports and Modes [P] on LC Detectors/Test Station/Station/PDA Detector Final Test TE82882-10	57	29	86	66.3
01	All Reports and Modes [P] on LC Detectors/Test Station/Station/PDA Detector Final Test TE82882-11	32	41	73	43.8
02	All Reports and Modes [P] on LC Detectors/Test Station/Station/PDA Detector Final Test TE82882-13	37	31	68	54.4
03	All Reports and Modes [P] on LC Detectors/Test Station/Station/PDA Detector Final Test TE82882-14	85	10	95	89.5