

Certificate of Analysis

ACQUITY UPLC/UV Qualification Standards Kit

Catalog Number: 700002642
Expiration Date: 06-OCT-2022
Lot Number: W06102013

Part Number	Lot Number	Description	Component	Concentration
700002642-1	W06102014	Solution 1	Caffeine	0.0100 mg/mL
700002642-2	W06102015	Solution 2	Caffeine	0.0300 mg/mL
700002642-3	W06102016	Solution 3	Caffeine	0.0400 mg/mL
700002642-4	W06102017	Solution 4	Caffeine	0.0500 mg/mL
700002642-5	W06102018	Solution 5	Caffeine	0.0800 mg/mL
700002642-6	W06102019	Solution 6	Caffeine	0.1000 mg/mL
700002642-7	W06102020	Solution 7	Caffeine	0.1600 mg/mL
700002642-8	W06102021	Solution 8	Caffeine	0.20 µg/mL
700002642-9	W06102022	Solution 9	Caffeine	4.000 mg/mL
700002642-10	W06102023	Solution 10	Thiourea Acetophenone	0.030 mg/mL 0.110 mg/mL
700002642-11	W06102024	Solution 11	2-Acetylfuran Acetanilide Acetophenone Propiophenone Butyl Paraben Benzophenone Valerophenone	4.00 µg/mL 4.00 µg/mL 4.00 µg/mL 4.00 µg/mL 4.00 µg/mL 4.00 µg/mL 4.00 µg/mL
700002642-12	W06102025	Solution 12	Blank	N/A

Solvent: Water/Acetonitrile – 90/10 v/v

This kit was prepared and certified at our ISO 9001 Registered facility in Golden, Colorado.

Balance used in manufacture of this standard is calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

The Safety Data Sheet (SDS) for this product is available at www.waters.com/ASRSDS.

Certificate Revision: Original



Kevin Thompson
Product Line Manager

Waters

THE SCIENCE OF WHAT'S POSSIBLE.™

34 Maple Street, Milford, MA 01757-3696, U.S.A.
Phone: 508-478-2000 Fax: 508-872-1990
www.waters.com

Supersedes/Date: 05/13/2015
Version: 05
Effective Date: 02-Apr-2019

ERA Doc #: S0081-132
Authored/Revised By:
Responsible Dept:
Page 1 of 1

CO #: 2042
Jimmy Felix
Waters