

QUALIFICATION SUMMARY REPORT


Project Name SQT_ACQUITY_TUV_12

System Name: Column Manager

Processing Node: Te82987-1

Database Service: Te82987-1

Qualification Tests	Performed	Pass	Fail	# of Pages
Test Injections	☒	☒	☐	3
System Precision	☒	☒	☐	3
Wavelength Accuracy	☒	☒	☐	3
UV Detector Linearity & Sensitivity	☒	☒	☐	5
Injector Linearity & Accuracy	☒	☒	☐	3
Detector Noise	☒	☒	☐	2
Flow Rate Linearity	☒	☒	☐	3
Carryover	☒	☒	☐	3
Gradient Performance	☒	☒	☐	3
Sample Organizer Temperature Accuracy	☐	☐	☐	
Column Heater / Cooler Temperature Accuracy	☐	☐	☐	
Sample Heater / Cooler Temperature Accuracy	☐	☐	☐	
Flow Rate Accuracy Pump A	☐	☐	☐	
Flow Rate Accuracy Pump B	☐	☐	☐	



QUALIFICATION SUMMARY REPORT

Project Name SQT_ACQUITY_TUV_12

System Name: Column Manager

Processing Node: Te82987-1

Database Service: Te82987-1

Certification

Comments

No comments.

NA

13 Apr 2021

The undersigned performer attests that the values recorded and calculated above are accurate and complete.

Performer

Signature

Date 13 Apr 2021

The undersigned reviewer accepts that the values recorded and calculated above are accurate and complete.

Reviewer

Signature

Date