

WATERS CORPORATION INSTRUMENT - FINAL TEST RECORD

Part Number: _____	Description: _____
Tested By: _____	Date: ____/____/____
Serial Number: _____	Control Number: _____
Software Revision: _____	Factory Order: _____

Electrical Safety

PASS ☐ FAIL ☐ see note #____ PASS ☐

FANS

PASS ☐ FAIL ☐ see note #____ PASS ☐

LED'S FUNCTIONALITY

PASS ☐ FAIL ☐ see note #____ PASS ☐

Linearity (% RSD) Actual: _____

PASS ☐ FAIL ☐ see note #____ PASS ☐

Wavelength Accuracy

255nm _____ 379 _____ 523 _____

PASS ☐ FAIL ☐ see note #____ PASS ☐

Final Test

PASS ☐ FAIL ☐ see note #____ PASS ☐

Instrument Result: PASS ☐ FAIL ☐ Signed _____ Date: ____/____/____

Notes: _____

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests noted above fail, a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing. The initial form must be attached and retained.

END OF DATA