

Waters Corporation - Instrument Final Test Record

Part Number: 186015070 Product: Acquity RI Detector

Operator: USIINKM Start Time: 04/07/2014, 09:19:06.000

Serial Number: D13URI015R

Firmware Rev. : 1.40 Build#: 44 (Mar 21 2013)

Name	LSL	Measurement	USL	Unit	Outcome
Acquity RI Download-Config					FAIL
Firmware Download					FAIL
Serial Number Set					PASS
Leak Detector Checked					PASS
Electrical Safety					PASS

Notes:

Signed: _____

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests noted above fail, a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing. The initial form must be attached and retained.