

Waters Corporation - Instrument Final Test Record

Part Number: 186015029

Product: Acquity FLR

Operator: SEOW PEI

Start Time: 01/07/2020
16:59:

Serial Number: A20UPF528G

Factory Order : WATS2001019

TMD# : 010759

Name	LSL	Measurement	USL	Unit	Outcome
Acquity FLR Final Test					PASS
Calibrate					PASS
PMTCal					PASS
Normalize					PASS
Restart					PASS
Rayleigh Wavelength	362.5	365.2	367.5	nM	PASS
Rayleigh Energy		6210327		EU	PASS
Gain	0	0.5	0.6		PASS
Raman Wavelength	413.5	414.4	418.5	nM	PASS
Raman Energy	100	46205	200000	EU	PASS
Ram-Ray Valley Energy	5000	6541	100000	EU	PASS
Optical Resolution	0	14.16	100		PASS
Stray Light Ratio	0	0.74	10		PASS
Signal/Noise Ratio	1000	1439.63	4000		PASS
PMT Coefficient	0	0.99983	2		PASS

Notes:

TE82760-5

END OF DATA

Signed: *Seow Pei*

Date: *07 JAN 20*

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests noted above fail, a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing. The initial form must be attached and retained.

Tuesday, January 07, 2020