

Waters Corporation - Test Record

Part Number: 186015029

Operator: HCTan

Serial Number: A20UPF526G

Product: Acquity FLR

Start Time: 01/08/2020, 10:32:

Factory Order : WATS2001019

TMD# : 010759

Name	LSL	Measurement	USL	Unit	Outcome
Calibrate					PASS
PMTCal					PASS
Normalize					PASS
Restart					PASS
Rayleigh Wavelength	362.5	365.2	367.5	nM	PASS
Rayleigh Energy		14709725		EU	PASS
Gain	0	0.5	0.6		PASS
Raman Wavelength	413.5	416	418.5	nM	PASS
Raman Energy	100	65534	200000	EU	PASS
Ram-Ray Valley Energy	5000	8645	100000	EU	PASS
Optical Resolution	0	13.19	100		PASS
Stray Light Ratio	0	0.45	10		PASS
Signal/Noise Ratio	1000	3289.12	4000		PASS
PMT Coefficient	0	0.999744	2		PASS

Notes: TES2760-7

END OF DATA

Signed: [Signature]

Date: 08 JAN 20

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests noted above fail, a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing. The initial form must be attached and retained.