

Part Number: <u>186002475</u>	Product: <u>2475 Detector</u>
Operator: <u>HCTan</u>	Start Time: <u>05/29/2019, 23:26:47</u>
Serial Number: <u>F1906020-03</u>	End Time: <u>05/29/2019, 23:45:03</u>
Test System: <u>2475/FLR Final Test System TE82760-7</u>	Order: <u>N/A</u>

Item	LSL	Measurement	USL	Unit	Outcome
Calibrate					PASS
ITCal					PASS
Normalize					PASS
Start					PASS
Wavelength	347.5	350.4	352.5	nM	PASS
Energy		14410824		EU	PASS
in		0.4			PASS
Wavelength	394.5	380.3	399.5	nM	FAIL
Energy	100	21227	200000	EU	PASS

2475 Detector Final Test Overall Result - FAIL

Signed:	Date:
---------	-------

ID OF DATA