



MULTIMETER CALIBRATION REPORT

Equipment : Multimeter Cal Date : 31 / Dec / 2020
Brand Name : FLUKE Next Cal Date : 31 / Dec / 2021
Model : FLUKE 115 Environment Cond. : N.A.
GES/Mfg Serial No. : 23372874 Equipment Used : Yokogawa Portable Calibrator
Cal Procedure : WI-QMS-044 S/No. 2422

Calibration Data

Calibrator Output	Accuracy		Measured Reading (before)	Measured Reading (after)
	Digital	Analog		
300 mV	297.9 ~ 302.1 mV	270.0 ~ 330.0 mV	300.1	N.A.
3 V	2.979 ~ 3.021 V	2.700 ~ 3.300 V	2.999	N.A.
30 V	29.79 ~ 30.21 V	27.00 ~ 33.00 V	29.99	N.A.
20 mA	19.86 ~ 21.14 mA	18.00 ~ 22.00 mA	NA	N.A.

Result :

Pass / Fail

Remarks :

Multimeter is tested ok within calibration range and Multimeter is not used to test current.

Calibrated by

Name/sign : Lee Kwang Quek

Designation : Senior Technician

Date : 31 DEC-2020

Reviewed by

Name/sign : Sheliyan Sathivelu

Designation : Senior QA Engineer

Date : 04 Jan 2021