

Waters Corporation - Test Record

Part Number: 186015085Product: H-Class FTN PlusOperator: 25705147Start Time: 06/03/2021, 11:51:Serial Number: D21FTP387GEnd Time: 06/03/2021, 14:17:Test System: TE0084050-18Order: 2100322Test Software: 1.03Location: GES/Venture,SG

TMD: TMD011012

TMD: TMD011014

Firmware: 1.71.395

Name	LSL	Measurement	USL	Unit	Outcome
Electrical Safety Test					PASS
Power-on Initialization					PASS
Fans					PASS
Firmware Download					PASS
Sample Compartment Door & Lights					PASS
Jumper Configuration		129.0	129.0		PASS
Test Event Inputs					PASS
Test Event Outputs					PASS
Test SPI Interface					PASS
Test Leak Detectors					PASS
Test Ecord Interface					PASS
Theta Drive Calibration					PASS
Tray Cal Offset	-180.0	33.89	180.0	Deg	PASS
R Axis Calibration					PASS
R Offset	-3.0	-0.02	3.0	mm	PASS
Z Axis Calibration					PASS
Z Offset	-3.0	-1.53	3.0	mm	PASS
Needle Seal Calibration					PASS
Needle Seal Position	48.0	56.46	68.0	mm	PASS
Needle Seal Force	48.0	61.81	68.0	lb/inch	PASS
Zp Offset	-3.0	1.29	3.0	mm	PASS
Sample Tray Select Switch					PASS
Check Theta Drive Belt					PASS
R/Theta Positioner Test					PASS
1st Move Error	-0.3	0.13	0.3		PASS
Zp Pierce	-20.0	1.0	20.0	Lost Steps	PASS
Z Pierce	-20.0	1.0	20.0	Lost Steps	PASS
R Lost Steps	-20.0	0.0	20.0	Lost Steps	PASS
Sample Cooler Test		-0.885	-0.6	Deg C/Min	PASS
Test Pressure Transducer	0.45	0.511	0.55	VDC	PASS
BOS Pump Test					PASS
Methanol Prime					PASS
Sample Syringe Decay	-15.0	3.8	15.0	PSI	PASS
Dynamic Leak Test		1.0	500.0	nL/min	PASS
Needlewash Prime					PASS
Check Acquity Stack					PASS
Test Preblank Injections		0.0	860.9	area	PASS
Carryover		0.0	0.06	ppm	PASS
Chemistry Precision Area	0.0	0.1221	0.5	%RSD	PASS
Chemistry Precision Height	0.0	0.2771	0.9	%RSD	PASS
Chemistry X-Intercept	-0.2	0.0059	0.2	ul	PASS
Chemistry Linearity (R^2)	0.999	0.999979	1.0		PASS

Methanol Purge
Air Purge
Park Inject Valve
Home Needle
Save Serial Number

PASS
PASS
PASS
PASS
PASS

CPSM Final Test Overall Result - PASS

Signed:	Date:
---------	-------

END OF DATA