

Waters Corporation - Test Record

Part Number: 186015032

Operator: 246154

Serial Number: G18UPD512A

Test System: Download-Configure System

Test Software: 1.4

Product: PDA Detector

Start Time: 07/10/2018, 10:30:

End Time: 07/10/2018, 10:37:

Order: NA

Location: Flextronics, SG


TMD: NA

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Firmware Download		Complete			PASS
Serial Number Set		Complete			PASS
Leak Detector Test		Passed			PASS
Electrical Safety		Passed			PASS

Acquity PDA Download-Config Overall Result - PASS

Signed:



Date:

10 JUL 2018

END OF DATA