

Waters Corporation - Test Record

Part Number: 186015033

Operator: 246154

Serial Number: G18UPL056A

Test System: Download-Configure System

Test Software: 1.4

Product: eLambda Detector

Start Time: 07/10/2018, 13:47:

End Time: 07/10/2018, 13:53:

Order: NA

Location: Flextronics, SG


TMD: NA

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Firmware Download		Complete			PASS
Serial Number Set		Complete			PASS
Leak Detector Test		Passed			PASS
Electrical Safety		Passed			PASS

**Acquity eLambda Download-Config Overall Result -
PASS**

Signed:



Date:

10 JUL 2018

END OF DATA