

Title: Installation Qualification Procedure

Document No: 730004135

Version No: 00

Page 1 of 2

IQ Test Name: Test Software and Instrument Firmware Version			Test Script No.: IQ-04	
Equipment No./ID: TE82969-1		Equipment Name: PDA/ eLambda Linearity Test System		
Installation Site: Flextronics: WPPL, Singapore				
Purpose: To verify that the installed firmware/software is correct per the applicable design/operational documents (TSP, ASP, MPR, BOM, and/or other Design documents). Test Execution: List all critical instruments/components and/or sub-systems that require software and/or firmware in the Description column. List the associated required software/firmware version, as defined in the applicable design/operational documents, in the Acceptance criteria column. Verify that the software/firmware version on the Test Equipment match the acceptance criteria and are identical. Enter N/A for any parameter or cell that is not used or applicable. Objective Evidence: Completed Test Script Data Sheet Identify the applicable design/operational document that defined the required software/firmware version(s) in the associated Installation Qualification Report.				
Description	Acceptance Criteria	As Found	Pass/Fail	Initial & Date
Firmware Version P/N: 186015034	Version: 1.70.6367	Version:		
Firmware Version P/N: 186015053	Version: 1.70.6367	Version:		
Comments: Firmware versions specified in ECR #9222852				
Overall Test Conclusion: ☐ Pass ☐ Fail				

Copy this Test Script as needed

