

Waters Corporation - Test Record

Part Number: 186015033

Operator: 842185

Serial Number: J18UPL246A

Test System: TE82969-1

Test Software: 2.1

Product: eLambda PDA

Start Time: 09/25/2018, 17:30:

End Time: 09/25/2018, 17:32:

Order: NA

Location: Flextronics,SG

TMD: NA

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
%RSD	0	4.263	5		PASS
Mean		0.09			PASS
Std Dev		0.004			PASS
Lambda		257.6			PASS
A/C 5mg		0.0929			PASS
A/C 10mg		0.0932			PASS
A/C 15mg		0.0933			PASS
A/C 20mg		0.0906			PASS
A/C 25mg		0.0875			PASS
A/C 30mg		0.0838			PASS

eLambda PDA Linearity Test Overall Result - PASS

Signed: L 842185

Date: 25 SEP 2018

END OF DATA