

WATERS CORPORATION 269X Alliance Heater Cooler Test System CALIBRATION and PREVENTIVE MAINTENANCE RECORD

Date: 25 Feb 19	Performed By: Mohammad Shamsul 246394 f	
Equipment I.D. PM 03300	Location: 5th floor, Production area (Kallang Place)	
Due Date: NA	Interval (Months): 12 months	Next Due Date: 25 Feb 20

Test Equipment Used	Serial Number/Identification Number	Calibration/PM Date
Digital Multimeter	Egid 0245	02 Jan 19
_____	End of data _____	_____
_____	_____	_____

SECTION 1: NI-9217 Calibration	RESULT	RESULT
Step 1.1: Calibration NI-2917	Calibration Status Valid ☐	PASS ☐ FAIL ☐ SEE NOTE ☒

SECTION 2: Computer Maintenance	RESULT
Step 2.1: Maintain Test Computer per 735003678, Generic PC Maintenance Procedure.	PERFORMED ☒ SEE NOTE ☐

SECTION 3: Vacuum and Cleaning	RESULT
Step 3.1: General Cleaning, inside and out.	PERFORMED ☒ SEE NOTE ☐

SECTION 4: Electrical	RESULT
Step 4.1: Inspect cables for wear or physical damage. Replace as needed.	PERFORMED ☒ SEE NOTE ☐
Step 4.2 Inspect connectors for wear or physical damage. Replace as needed.	PERFORMED ☒ SEE NOTE ☐

SECTION 5: Mechanical	RESULT
Step 5.1: Inspect packaging physical damage/deformation.	PERFORMED ☒ SEE NOTE ☐
Step 5.2: Inspect packaging or sharp edges	PERFORMED ☒ SEE NOTE ☐
Step 5.3: Check inside and around tester enclosure for loose hardware.	PERFORMED ☒ SEE NOTE ☐

SECTION 6: Functional Check	RESULT
Step 6.1: Run the test using a known good UUT. Verify a passing result is achieved.	PASS ☒ FAIL ☐ SEE NOTE ☐

NOTES

Step 1.1 . Equipment have yet to be send for calibration.

_____ End of data _____