

Waters Corporation - Test Record


Part Number: <u>700012245</u>	Product: <u>eLambda Optics Bench On-Semi</u>
Operator: <u>246154</u>	Start Time: <u>07/04/2019, 16:24:</u>
Serial Number: <u>22K4B002</u>	End Time: <u>07/04/2019, 19:37:</u>
Test System: <u>TE82882-10</u>	Order: <u>N/A</u>
Test Software: <u>4.2</u>	Location: <u>Flextronics, SG</u>

TMD: N/A

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Baseline Temperature (Audit)	20	25.78	40	Degrees C	PASS
Hot Baseline (Audit)	20	26.23	40	Degrees C	PASS
F1 alone (Audit)	20	27.42	40	Degrees C	PASS
F1 and F2 (Audit)	20	25.96	40	Degrees C	PASS
Electrical Safety		True			PASS
Lambda1	178	185.6704	188	nM	PASS
Lambda512	801.5	804.3227	811.5	nM	PASS
Shunt/Cell energy 5ms	80000	80062	250000	counts	PASS
Total Runs	0	10	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0	4.52	7	uAu	PASS
Final Drift	-1000	885	1000	uAu	PASS

eLambda Optics Bench Final Test Overall Result - PASS

Signed: 246154 	Date: 04 JUL 2019
---	--------------------------

END OF DATA