

Waters Corporation

INSTRUMENT - FINAL TEST RECORD

Part Number: 186015043 Description: ACQUITY UPLC CM-A
 Tested By: 25704791 Date: 4/8/2019
 Serial Number: C19CMP535G TMD Number: N/A
 Firmware Revision: 1.69.154 Factory Order: WATS1904037

Electrical Safety Test	Pass ☒	Fail ☐	see note# _____	Pass ☐
Fan Test	Pass ☒	Fail ☐	see note# _____	Pass ☐
Ethernet Port Test	Pass ☒	Fail ☐	see note# _____	Pass ☐
LED Test	Pass ☒	Fail ☐	see note# _____	Pass ☐
Column LED Test	Pass ☒	Fail ☐	see note# _____	Pass ☐
Door Test	Pass ☒	Fail ☐	see note# _____	Pass ☐
Ecord 3	Pass ☒	Fail ☐	see note# _____	Pass ☐
Ecord 4	Pass ☒	Fail ☐	see note# _____	Pass ☐
Ecord 5	Pass ☒	Fail ☐	see note# _____	Pass ☐
Ecord 6	Pass ☒	Fail ☐	see note# _____	Pass ☐
Set Serial Number	Pass ☒	Fail ☐	see note# _____	Pass ☐
Fuses	Pass ☒	Fail ☐	see note# _____	Pass ☐
Leak Detector Interface Test	Pass ☒	Fail ☐	see note# _____	Pass ☐
Pre-Heater Interface TC-CM-APH1	Pass ☒	Fail ☐	see note# _____	Pass ☐
Pre-Heater Interface TC-CM-APH2	Pass ☒	Fail ☐	see note# _____	Pass ☐
Pre-Heater Interface TC-CM-APH3	Pass ☒	Fail ☐	see note# _____	Pass ☐
Pre-Heater Interface TC-CM-APH4	Pass ☒	Fail ☐	see note# _____	Pass ☐
TOP Auxillary Interface	Pass ☒	Fail ☐	see note# _____	Pass ☐
BOTTOM Auxillary Interface	Pass ☒	Fail ☐	see note# _____	Pass ☐
Heater Cooler	Pass ☒	Fail ☐	see note# _____	Pass ☐
Heater Cooler	Pass ☒	Fail ☐	see note# _____	Pass ☐
Left Valve	Pass ☒	Fail ☐	see note# _____	Pass ☐
Right Valve	Pass ☒	Fail ☐	see note# _____	Pass ☐

Instrument Result:

Pass ☒

Fail ☐

Signed:

[Signature]
END OF DAY

Date:

08 APR 19

Notes:

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests noted above fail; a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing; the initial form must be attached and retained.