

Waters Corporation

INSTRUMENT - FINAL TEST RECORD

Part Number: 186015038 Description: ACQUITY Active Column Manager

Tested By: 25704791 Date: 4/10/2019

Serial Number: D19CMP550G TMD Number: N/A

Firmware Revision: 1.60.136 Factory Order: WATS1904063

Electrical Safety Test	Pass ☒	Fail ☐	see note#	Pass ☐
Fan Test	Pass ☒	Fail ☐	see note#	Pass ☐
Ethernet Port Test	Pass ☒	Fail ☐	see note#	Pass ☐
LED Test	Pass ☒	Fail ☐	see note#	Pass ☐
Column LED Test	Pass ☒	Fail ☐	see note#	Pass ☐
Door Test	Pass ☒	Fail ☐	see note#	Pass ☐
Ecord 3	Pass ☒	Fail ☐	see note#	Pass ☐
Ecord 4	Pass ☒	Fail ☐	see note#	Pass ☐
Ecord 5	Pass ☒	Fail ☐	see note#	Pass ☐
Ecord 6	Pass ☒	Fail ☐	see note#	Pass ☐
Set Serial Number	Pass ☒	Fail ☐	see note#	Pass ☐
Fuses	Pass ☒	Fail ☐	see note#	Pass ☐
Leak Detector Interface Test	Pass ☒	Fail ☐	see note#	Pass ☐
Pre-Heater Interface TC-CM-APH1	Pass ☒	Fail ☐	see note#	Pass ☐
Pre-Heater Interface TC-CM-APH2	Pass ☒	Fail ☐	see note#	Pass ☐
Pre-Heater Interface TC-CM-APH3	Pass ☒	Fail ☐	see note#	Pass ☐
Pre-Heater Interface TC-CM-APH4	Pass ☒	Fail ☐	see note#	Pass ☐
TOP Auxillary Interface	Pass ☒	Fail ☐	see note#	Pass ☐
BOTTOM Auxillary Interface	Pass ☒	Fail ☐	see note#	Pass ☐
Heater Cooler	Pass ☒	Fail ☐	see note#	Pass ☐
Heater Cooler	Pass ☒	Fail ☐	see note#	Pass ☐
Left Valve	Pass ☒	Fail ☐	see note#	Pass ☐
Right Valve	Pass ☒	Fail ☐	see note#	Pass ☐

Instrument Result:

Pass ☒

Fail ☐

Signed:

[Signature]

Date:

10 APR 19

Notes:

END OF DAY

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests noted above fail; a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing; the initial form must be attached and retained.