

Waters Corporation - Test Record

Part Number: <u>186015049</u>	Product: <u>Acquity UPLC CM-AUX</u>
Operator: <u>25704791</u>	Start Time: <u>04/08/2019, 16:26:</u>
Serial Number: <u>D19CMX722G</u>	End Time: <u>04/08/2019, 17:18:</u>
Test System: <u>Acquity CMA AUX Test TE 83026-1</u>	

Name	LSL	Measurement	USL	Unit	Outcome
Electrical Safety Test					PASS
AUX ID					PASS
LED Test					PASS
Fan Test					PASS
Column LED Test					PASS
Door Test					PASS
Ecord 1					PASS
Ecord 2					PASS
Leak Detector Interface Test					PASS
Pre-Heater Interface TC-CHC1-APH1					PASS
Pre-Heater Interface TC-CHC1-APH2					PASS
Pre-Heater Interface TC-CHC1-APH1					PASS
Pre-Heater Interface TC-CHC1-APH2					PASS
Heater Cooler					PASS
Heater Cooler					PASS

Acquity CM AUX Test Overall Result - PASS

Signed:	Date:
---------	-------

END OF DATA