

# Waters Corporation

## INSTRUMENT - FINAL TEST RECORD

Part Number: 186015049 Description: Acquity UPLC CM-AUX  
Tested By: 25704791 Date: 4/8/2019  
Serial Number: D19CMX722G TMD Number: TMD: 010641, TMD: 010643  
Firmware Revision: N/A Factory Order: WATS1904065

|                                   |                                          |                               |           |                               |
|-----------------------------------|------------------------------------------|-------------------------------|-----------|-------------------------------|
| Electrical Safety Test            | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| AUX ID                            | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| LED Test                          | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Fan Test                          | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Column LED Test                   | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Door Test                         | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Ecord 1                           | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Ecord 2                           | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Leak Detector Interface Test      | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Pre-Heater Interface TC-CHC1-APH1 | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Pre-Heater Interface TC-CHC1-APH2 | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Pre-Heater Interface TC-CHC1-APH1 | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Pre-Heater Interface TC-CHC1-APH2 | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Heater Cooler                     | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |
| Heater Cooler                     | Pass <input checked="" type="checkbox"/> | Fail <input type="checkbox"/> | see note# | Pass <input type="checkbox"/> |

Instrument Result:

Pass ☒

Fail ☐

Signed:

*[Signature]*

Date:

08 APR 19

Notes:

Console test pass

END OF DATA

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests noted above fail; a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing; the initial form must be attached and retained.