

Waters Corporation - Test Record

Part Number: <u>186018007</u>	Product: <u>Acquity Premier eLambda PDA</u>
Operator: <u>1093658</u>	Start Time: <u>08/19/2020, 17:24:</u>
Serial Number: <u>H20PPD002A</u>	End Time: <u>08/19/2020, 18:10:</u>
Test System: <u>TE82882-10</u>	Order: <u>NA</u>
Test Software: <u>4.2</u>	Location: <u>Flextronics, SG</u>

TMD: NA

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Baseline Temperature (Audit)	20	24.46	40	Degrees C	PASS
Hot Baseline (Audit)	20	25.19	40	Degrees C	PASS
F1 alone (Audit)	20	26.61	40	Degrees C	PASS
F1 and F2 (Audit)	20	24.37	40	Degrees C	PASS
Electrical Safety		True			PASS
Lambda1	178	186.3663	188	nM	PASS
Lambda512	801.5	804.754	811.5	nM	PASS
Shunt/Cell energy 5ms	80000	104309	250000	counts	PASS
Total Runs	0	2	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0	5.4	7	uAu	PASS
Final Drift	-1000	704	1000	uAu	PASS

Acquity Premier eLambda PDA Final Test Overall Result -
PASS

Signed: <i>L. 1093658</i>	Date: <i>20 Aug 2020</i>
----------------------------------	---------------------------------

END OF DATA

Waters Corporation - Test Record

Part Number: 186018007

Operator: 1093658

Serial Number: H20PPD002A

Test System: Download-Configure System

Test Software: 1.4

Product: eLambda Detector

Start Time: 08/19/2020, 16:50:

End Time: 08/19/2020, 16:56:

Order: NA

Location: Flextronics, SG

TMD: NA

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Firmware Download		Complete			PASS
Serial Number Set		Complete			PASS
Leak Detector Test		Passed			PASS
Electrical Safety		Passed			PASS

Acquity Premier eLambda PDA Download-Config Overall Result - PASS

Signed: L. 1093658

Date: 19 AUG 2020

END OF DATA