

Waters Corporation - Test Record

Part Number: <u>186018007</u>	Product: <u>Acquity Premier eLambda PDA</u>
Operator: <u>246154</u>	Start Time: <u>08/28/2020, 10:06:</u>
Serial Number: <u>H20PPD016A</u>	End Time: <u>08/28/2020, 11:11:</u>
Test System: <u>TE82882-13</u>	Order: <u>N/A</u>
Test Software: <u>4.2</u>	Location: <u>Flextronics, SG</u>

TMD: N/A

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Baseline Temperature (Audit)	20	22.3	40	Degrees C	PASS
Hot Baseline (Audit)	20	22.59	40	Degrees C	PASS
F1 alone (Audit)	20	23.46	40	Degrees C	PASS
F1 and F2 (Audit)	20	22.63	40	Degrees C	PASS
Electrical Safety		True			PASS
Lambda1	178	186.2925	188	nM	PASS
Lambda512	801.5	804.5917	811.5	nM	PASS
Shunt/Cell energy 5ms	80000	112626	250000	counts	PASS
Total Runs	0	3	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0	5.56	7	uAu	PASS
Final Drift	-1000	57	1000	uAu	PASS

Acquity Premier eLambda PDA Final Test Overall Result
- PASS

Signed: W 1093658

Date: 28 Aug 2020

END OF DATA