

Waters Corporation - Test Record

Part Number: <u>186018007</u>	Product: <u>Acquity Premier eLambda PDA</u>
Operator: <u>1093658</u>	Start Time: <u>09/03/2020, 16:22:</u>
Serial Number: <u>H20PPD020A</u>	End Time: <u>09/03/2020, 18:20:</u>
Test System: <u>TE82882-14</u>	Order: <u>NA</u>
Test Software: <u>4.2</u>	Location: <u>Flextronics, SG</u>

TMD: NA

Firmware: 1.70 Build#: 6367 (Jan 25 2018)

Name	LSL	Measurement	USL	Unit	Outcome
Baseline Temperature (Audit)	20	23	40	Degrees C	PASS
Hot Baseline (Audit)	20	23.56	40	Degrees C	PASS
F1 alone (Audit)	20	24.57	40	Degrees C	PASS
F1 and F2 (Audit)	20	23.06	40	Degrees C	PASS
Electrical Safety		True			PASS
Lambda1	178	185.987	188	nM	PASS
Lambda512	801.5	803.6171	811.5	nM	PASS
Shunt/Cell energy 5ms	80000	113695	250000	counts	PASS
Total Runs	0	6	24		PASS
Noise Runs	0	1	5		PASS
Final Noise	0	5.84	7	uAu	PASS
Final Drift	-1000	-812	1000	uAu	PASS

**Acquity Premier eLambda PDA Final Test Overall Result
- PASS**

Signed: ①fmc 246231

Date: 03 SEP 2020

END OF DATA