

CALTEK PTE. LTD.

23 Tagore Lane, #04-08/09/10/11 Tagore 23 Warehouse, Singapore 787601
Tel: (65) 64520300 Fax: (65) 64520500
Email: info@caltexgroup.com Website: www.caltexgroup.com

CALIBRATION CERTIFICATE

CERTIFICATE NUMBER : CTS-111-1014-20	JOB NUMBER : CTJ19-8721
RECEIVED DATE : 30-Dec-19	ISSUE DATE : 08-Jan-20

INSTRUMENT : THERMOHYGROGRAPH	Ambient Temperature : (23 ± 5) °C
MANUFACTURER : HISAMATSU MFG. CO.	Relative Humidity : (55 ±10) % r.h.
MODEL NO : TN-2500	Date of Calibration : 07-Jan-20
PART NO : -	Recommended Due Date : 07-Jan-21
SERIAL NO : 5650	Instrument Code : CT3260
CUSTOMER : FLEXTRONICS MANUFACTURING (SINGAPORE) PTE LTD.	Range : -
	Asset/Tag/Id/Code : 0418
	Page No : 1 of 2
	Status : As Found
	JR No : JR-17388

The described instrument has been calibrated at Caltex laboratory under the ambient conditions stated above.

This certificate provides traceability of measurement to the International System of Units (SI) and/or to units of measurement realised at the National Metrology Centre (NMC), Singapore or other recognized national metrology institutes.

The calibration method was carried out according to In-house Technical Calibration Procedure CTTM-T10-2007, as guide.

S.No.	REFERENCE INSTRUMENT(S)	SERIAL NO	DUE DATE
1.	Std.Hygrometer c/w probe	60813824/20055608	02-May-20

RESULTS OF CALIBRATION :

- The expanded uncertainty of measurement associated with the calibration is estimated at a level of confidence of approximately 95 % with a coverage factor of k=2.00.
- The results of calibration are given on the attached calibration data sheet(s).
- The user should determine the suitability of the instrument for its intended use.
- No Adjustments done.

REVIEWED & ACCEPTED

SIGN: CHANG MEL LAN

DATE: 09 JAN 2020

REMARKS: FOLD 0418

Yepthimich 10 JAN 20

Calibrated by :

Saathyish Vejayandran

EMP ID : 1107


Approved by :

Siva

EMP ID : NA

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing Laboratory.
The results reported herein have been performed in accordance with the terms of accreditation under the Singapore Accreditation Council.





CALTEK PTE. LTD.

23 Tagore Lane, #04-08/09/10/11 Tagore 23 Warehouse, Singapore 787601
Tel: (65) 64520300 Fax: (65) 64520500
Email: info@caltekgroup.com Website: www.caltekgroup.com

CALIBRATION CERTIFICATE

CERTIFICATE NUMBER : CTS-111-1014-20

JOB NUMBER : CTJ19-8721

ISSUE DATE : 08-Jan-20

PAGE NUMBER : 2 of 2

Calibration Results: RELATIVE HUMIDITY MEASUREMENT @ FIXED TEMPERATURE OF 23°C

Mean Reference Reading	Mean Instrument Reading - Before Adjustment	Error	Expanded Uncertainty
%rh	%rh	%rh	%rh
24.7	25.5	0.8	2.5
49.8	50.0	0.2	2.5
84.3	85.5	1.2	3.0

Calibration Results: TEMPERATURE MEASUREMENT @ NOMINAL RELATIVE HUMIDITY OF 50%r.h.

Mean Reference Reading	Mean Instrument Reading - Before Adjustment	Error	Expanded Uncertainty
°C	°C	°C	°C
17.8	18.2	0.4	0.6
23.9	24.2	0.3	0.6
29.8	30.4	0.6	0.6

Calibrated by :

Saathyish Vejayandran

EMP ID : 1107

Approved by :

Siva

EMP ID : NA

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing Laboratory.
The results reported herein have been performed in accordance with the terms of accreditation under the Singapore Accreditation Council.

