

WATERS CORPORATION - TEST RECORD

Part Number:	186018006	Product:	ACQUITY Premier TUV DETECTOR
Operator:	Ho Teck Lik	Start Date/Time:	18 Aug 2020 1013
Serial Number:	H20PTV001A	End Date/Time:	18 Aug 2020 1559
Test System:	TUV Download / pretest station	Order:	NA
Test Software:	NA	Location:	flextronics-sg

TMD: NA

Firmware: 1.70.2648

Test Name	LSL	Measurement	USL	Unit	Outcome		
Visual Inspection					PASS ☒	FAIL ☐	See Event Log ☐
Electrical Safety					PASS ☒	FAIL ☐	See Event Log ☐
Calibration					PASS ☒	FAIL ☐	See Event Log ☐
Fans					PASS ☒	FAIL ☐	See Event Log ☐
Lamp ID					PASS ☒	FAIL ☐	See Event Log ☐
Flowcell ID					PASS ☒	FAIL ☐	See Event Log ☐

Note: Fill in 'N/A' if not applicable.

Instrument Overall Result - PASS ☒ FAIL ☐

Signed: C. 1093658	Date: 18 Aug 2020
--------------------	-------------------

END OF DATA

Document No.: _____
Iteration No.: _____
Attachment: _____ Page ____ of ____