

5.0 TRAINING

The operator of the TE82964-1, the TUV DOWNLOAD / PRETEST STATION was trained according to the Test Specification and Procedure, ACQUITY PREMIER TUV DETECTOR 735005108, Rev A. The executor of this qualification was trained to the Installation Qualification Protocol, 730004135 Ver. 03. Both Training records are provided as Attachments 7.1 and 7.2, respectively.

6.0 QUALIFICATION Results

6.1 Test Script IQ-01: Associated Documentation Verification was performed according to Installation Qualification Protocol 730004135, Ver. 03. The completed Test Script Data Sheet is provided as Attachment 7.3.

6.1.1 The Test Specification and Procedure, ACQUITY PREMIER TUV DETECTOR 735005108, Rev A was confirmed to be in SAP with a status of Ready. A screen capture from SAP is provided as Attachment 7.4.

6.2 Test Script IQ-02: Site Verification was not required to be performed as part of this test equipment qualification according to TE82964-05AUG20-VAF.

6.3 Test Script IQ-03: Site Verification was not required to be performed as part of this test equipment qualification according to TE82964-05AUG20-VAF.

6.4 Test Script IQ-04: Site Verification was not required to be performed as part of this test equipment qualification according to TE82964-05AUG20-VAF.

6.5 Test Script IQ-05: Site Verification was not required to be performed as part of this test equipment qualification according to TE82964-05AUG20-VAF.

6.6 Test Script IQ-06: Functional Verification was performed according to Installation Qualification Protocol 730004135, Ver. 03. The completed Test Script Data Sheet is provided as Attachment 7.5.

6.6.1 One (1) UUTs was tested using the TE82964-1, the TUV DOWNLOAD / PRETEST STATION with all tests passing. The screenshot of passing result is provided as Attachment 7.6.

6.7 This table provides a summary of each test script as performed according to 730004135, Ver. 03 along with the attachment location of all objective evidence. The table also indicates any deviation taken from the scripts during qualification.

Script or Section	Pass/Fail	Objective Evidence	Deviations
IQ-01	Pass	Attachment 7.3 and 7.4	None
IQ-06	Pass	Attachments 7.5 and 7.6	None

7.0 ATTACHMENTS

7.1 Completed Operator Training Record 735005108, Rev A

7.2 Completed Executor Training Record 730004135, Ver. 03

7.3 Completed Test Script IQ-01: Associated Document Verification

7.4 Objective Evidence IQ-01: Screenshot of SAP

7.5 Completed Test Script IQ-06: Functional Verification

7.6 Objective Evidence IQ-06: Screenshot of passing result on UUT

8.0 DEVIATION FROM THE QUALIFICATION PROTOCOL

There were no deviations generated during the execution of the Installation Qualification Protocol 730004135, Ver. 03.