

Waters Corporation

INSTRUMENT FINAL TEST RECORD

Part Number:	186002449	Description:	Pump Control Module II
Tested By:	Alex EE	Date:	4/1/2021 11:13:18 AM
Serial Number:	M18PC2543G	Control Number:	N/A
Software Revision:	4.00 Build#: 12 (Oct 15 2008)	Factory Order:	1234567

- 1) Hi - Pot / Ground Continuity Test PASS ☒ FAIL ☐ see note # _____ PASS ☐
- 2) Rear Panel I/O Test PASS ☒ FAIL ☐ see note # _____ PASS ☐
- 3) Pumps A, B, & C Status Test
(-12 VDC & +12 VDC) PASS ☒ FAIL ☐ see note # _____ PASS ☐
- 4) Pumps A, B, & C Pressure Test
(-10 VDC - 0 VDC) PASS ☒ FAIL ☐ see note # _____ PASS ☐
- 5) Pumps A, B, & C Control Test
(0 - 2133 HZ TTL) PASS ☒ FAIL ☐ see note # _____ PASS ☐

Instrument Result: PASS ☒ FAIL ☐ Signed: 

Date: 01/APR/2021

Notes:

N/A

This Test Record and any data contained within it are proprietary to Waters Corporation and must be retained by Waters for the time period prescribed in its Quality Records Retention Policy. If any of the individual tests noted above fail; a note must be referenced that details the work performed to correct that specific deficiency. An overall failure indicates a special circumstance in which case a complete retest is to be done. A second form is required to document that testing; the initial form must be attached and retained.